

STOCK Tg(tetNZL)2Bjd/J

Stock No: 003274

Protocol 24893: Standard PCR Assay - Luciferase: Tg(tetNZL)2Bjd, Tg(tetL)1Bjd, Tg(Fos-luc)1Rnd

Version 1.2

Notes

This assay will NOT distinguish hemizygous from homozygous transgenic animals.

The genotyping protocol(s) presented here have been optimized for reagents and conditions used by The Jackson Laboratory (JAX). To genotype animals, JAX recommends researchers validate the assay independently upon receipt of animals into their facility. Reaction cycling temperature and times may require additional optimization based on the specific genotyping reagents used.

Expected Results

Transgene = 465 bp

Internal positive control = 324 bp

Transgene Tm = 82 Area = 11 Tm = 85.5 Area = 30

IPC Tm = 82 Area = 50

JAX Protocol

Protocol Primers

PRIMER	5' LABEL	SEQUENCE 5' → 3'	3' LABEL	PRIMER TYPE	REACTION	NOTE
oIMR0960		TCC TCT GAC ACA TAA TTC GCC		Transgene Forward	A	
oIMR0961		GTT GTT CCA TTC CAT CAC GG		Transgene Reverse	A	
oIMR7338		CTA GGC CAC AGA ATT GAA AGA TCT		Internal Positive Control Forward	B	
oIMR7339		GTA GGT GGA AAT TCT AGC ATC ATC C		Internal Positive Control Reverse	B	

Reaction A

COMPONENT	FINAL CONCENTRATION
ddH2O	
Kapa 2G HS buffer	1.30 X
MgCl ₂	2.60 mM
dNTPS-kapa	0.26 mM
oIMR0960	0.50 uM
oIMR0961	0.50 uM
Glycerol	6.50 %
Dye	1.00 X
Kapa 2G HS taq polym	0.03 U/ul
DNA	

Cycling

STEP	TEMP °C	TIME	NOTE
1	94.0	--	
2	94.0	--	
3	65.0	--	-0.5 C per cycle decrease
4	68.0	--	
5		--	repeat steps 2-4 for 10 cycles (Touchdown)
6	94.0	--	
7	60.0	--	
8	72.0	--	
9		--	repeat steps 6-8 for 28 cycles
10	72.0	--	
11	10.0	--	hold

JAX uses a very high speed Taq (~1000 bp/sec), use cycling times recommended for your reagents.

JAX uses a 'touchdown' cycling protocol and therefore has not calculated the optimal annealing temperature for each set of primers.

Reaction B

COMPONENT	FINAL CONCENTRATION
ddH2O	
Kapa 2G HS buffer	1.30 X
MgCl ₂	2.60 mM
dNTPS-kapa	0.26 mM
oIMR7338	0.50 uM
oIMR7339	0.50 uM
Glycerol	6.50 %
Dye	1.00 X
Kapa 2G HS taq polym	0.03 U/ul
DNA	

Cycling

STEP	TEMP °C	TIME	NOTE
1	94.0	--	
2	94.0	--	
3	65.0	--	-0.5 C per cycle decrease
4	68.0	--	
5		--	repeat steps 2-4 for 10 cycles (Touchdown)
6	94.0	--	
7	60.0	--	
8	72.0	--	
9		--	repeat steps 6-8 for 28 cycles
10	72.0	--	
11	10.0	--	hold

JAX uses a very high speed Taq (~1000 bp/sec), use cycling times recommended for your reagents.

JAX uses a 'touchdown' cycling protocol and therefore has not calculated the optimal annealing temperature for each set of primers.

